

LAB ON LOCALE™

Portable Gas Analysis System



Customer Name	MERIDIANO 87 SA DE CV
Test Date / Technician	6/29/2021 2:02:03 PM / Maria Silva
Lot Number	
Compressor Model	Compressor
Compressor Serial Number	000
Sample Pressure	100psi
Test Standard	CGA E (G-7.1, 2011)

Equipment that was filled by source:

Analytes	Sample	Standard	Accuracy
Oxygen (%)	21.5	20 - 22	± 1%
Carbon Dioxide (ppm)	573.6	1000	± 2%
Carbon Monoxide (ppm)	<1	10	± 2%
Methane (ppm)	<1	25	± 2%
Dew Point (°F)	-65.8	-65	± 2%
Oil & Particles (mg/m3)	Pass	5	± 2%
Odor*	None	None	NA

*Performed by Customer

Results

Sample Meets Standard

Test performed by Lawrence Factor, Inc. for Lawrence Factor Mexico - LOL000979

IMPORTANT: This air/gas or substance was tested only against a specified standard and may contain undetected items which are beyond the purpose or scope of this analysis. More extensive testing can be conducted upon request. This does not guarantee the condition nor safe application of the analyzed air/gas substance. Results reported relate only to the items tested. This report shall not be reproduced except in full, without the written consent of Lawrence Factor, Inc. Methods Used: LF-501. Technologies used: IR, electrochemical, color sensor, organoleptic. Estimated Measurements of Uncertainty values for the reported analytes are available upon request. This report / certificate is not issued for an unbounded duration.

This report contains information which is confidential and may be legally privileged. The information is intended only for the individual or entity named on this report. If you are not the intended recipient, be aware that any disclosure, copying, distribution or use of this information is prohibited. If you received this report in error, please notify us by phone immediately so that we can arrange for retrieval of the original report.

Supplement to Test Report



APLAC® IAAC

ilac.

Technician / Maria Silva

Endorsement / President

X-zam Laboratories